

Notice of Allowability

Application No.

10/798,794

Examiner

Jon-Eric C. Morales

Applicant(s)

HARTLEY ET AL.

Art Unit

3766

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 9/14/2007.
2. ☒ The allowed claim(s) is/are 1-36.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Clara Davis on 9/27/2007.

2. The application has been amended as follows:

Claim 20 (Currently Amended):

A system for providing disordered breathing therapy to a patient, comprising:
a circuit configured to detect cardiac beats and measure cardiac intervals between cardiac beats;
a controller coupled to the circuit and configured to compute a first indicated pacing interval based at least on a measured cardiac interval duration and a previous value of the first indicated pacing interval; and
a cardiac pacing circuit coupled to the controller and configured to provide cardiac pacing to mitigate disordered breathing based on the first indicated pacing interval.

3. Claims 1, 20 and 31 are allowed over prior art of record.

The following is an examiner's statement of reasons for allowance: In regards to independent claim 1, the method of delivering disordered breathing therapy including measuring cardiac intervals between cardiac beats and computing a first indicated pacing interval based at least on a measured cardiac interval duration and a previous value of the first indicated pacing interval has not been disclosed or suggested in the prior art when combined with the rest of the limitations of claim 1.

Art Unit: 3766

With respect to claims 20, the system for providing disordered breathing therapy including a circuit configured to detect cardiac beats and measure cardiac intervals between cardiac beats and a controller coupled to the circuit and configured to compute a first indicated pacing interval based at least on a measured cardiac interval duration and a previous value of the first indicated pacing interval has not been disclosed or suggested in the prior art when combined with the rest of the limitations of claim 20.

Concerning claim 31, the system for providing disordered breathing therapy including means for measuring cardiac intervals between cardiac beats and means for computing a first indicated pacing interval based at least on a cardiac interval duration and a previous value of the first indicated pacing interval has not been disclosed or suggested in the prior art when combined with the rest of the limitations of claim 31.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jon-Eric C. Morales whose telephone number is 571-272-3107. The examiner can normally be reached on Monday through Friday from 8am - 5pm EST.

Art Unit: 3766


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Carl H. Layno can be reached on 571-272-4949. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Jon-Eric Morales
/Jon-Eric Morales/
Patent Examiner
Art Unit 3766

JEM

Carl Layno
Supervisory Primary Examiner
Art Unit 3766


CARL LAYNO
PRIMARY EXAMINER